

Alkalinity

Std. Method 2320 B-Titration

-alternatively-

The old EPA 310.2-colorimetric

Hardness

Std. Method 2340 B-Calculation

-alternatively-

Std. Method 2340 C-EDTA titration;

EPA 130.1, colorimetric,

EPA 130.2, (same as SM 2340 C)

pH

Std. Method 4500 H and EPA 150.1 are both the same, electrometric (pH electrode).

Free and total Chlorine

Std. Method 4500-Cl-G-DPD Colorimetric or **Std. Method 4500-Cl D-Amperometric Titration**

TDS

Std. Method 2540 C. Equivalent to the old EPA 160.1 method.

NOTE - the older EPA methods are sometimes used in field-testing, many of these older EPA methods are more portable.