

Task Group on Design and Construction Requirements
49i173Mr1 Straw Ballot
May 14, 2024

This document is the property of NSF International (NSF) and is for NSF Committee purpose only. Unless given prior approval from NSF, it **shall not** be reproduced, circulated, or quoted, in whole or in part, outside of NSF.

Purpose

The purpose of this ballot is to affirm revised and new language related to the many design and construction requirements throughout Standard 49.

Background

Issue paper **BSC-2022-03 – Design and Construction Requirements** highlighted that the design and construction requirements in sections 4 and 5 haven't seen much evaluation in many years. The issue proponent suggested the establishment of task group (TG) to discuss these sections. This issue was presented to the JC during the 2022 Face-to-Face meeting at which time the TG was motioned into existence. Since that time, the TG has met four times and conducted a straw ballot with feedback in November 2023.

The straw ballot was quite extensive with good feedback, and the TG discussed breaking up the proposal into several subsection ballots to make the work more digestible. During the most recent meeting (April 22, 2024) the issue proponent presented 18 different subsections and the group agreed to send each to straw ballot first with this TG. Any subsections receiving no further revision proposals will then be sent to the full JC for approval ballot. In this manner, some if not all of the proposed revisions will be ready for the next publication and the entire ballot will not be held up by a few subsections.

This ballot language reflects **i173M – Interlock Alarm Activation**

The **grey highlighted** portions of the language are proposed additions to the language of the standard.
The **strikeout** portions of the language are proposed deletions to the language of the standard.

An **affirmative (yes) vote** on this straw ballot means you agree with the revised language as submitted.

A **negative (no) vote** on this straw ballot means you disagree with the revised language as submitted. A negative vote must include an explanation of why you disagree with the revised draft.

Not for publication. This document is part of the NSF standard development process. This draft text is for circulation for review and/or approval by a NSF Standards Committee and has not been published or otherwise officially adopted. All rights reserved. This document may be reproduced for informational purposes only.

[Note – the recommended changes to the standard which include the current text of the relevant section(s) indicate deletions by use of ~~strikeout~~ and additions by **grey highlighting**. Rationale statements are in *red italics* and only used to add clarity; these statements will NOT be in the finished publication.]

NSF/ANSI International Standard for Biosafety Cabinetry —

Biosafety Cabinetry: Design, Construction, Performance, and Field Certification

-

5 Design and Construction

-

5.25.2 Internal cabinet supply / exhaust fan interlock alarm

When a cabinet contains both an internal downflow and exhaust fan, they shall be interlocked so that the downflow fan shuts off whenever the exhaust fan fails. An audible and visual alarm shall signal the failure. If the downflow fan fails, the exhaust fan shall continue to operate, and an audible and visual alarm shall signal the failure. Audible and visual alarms shall activate within 15 seconds of the alarm condition initiating.

NSF explanation: This applies the same time requirement specified for the type B exhaust and canopy alarms.